

Special 510(k) Premarket Notification
GE Medical Systems - LOGIQ 5 Ultrasound System
December 11, 2001

JAN 1 0 2002

Attachment B:
Summary of Safety and Effectiveness
Prepared in accordance with 21 CFR Part 807.92(c).



GE Medical Systems

General Electric Company
P.O. Box 414, Milwaukee, WI 53201

Section a):

1. **Submitter:** GE Medical Systems
PO Box 414
Milwaukee, WI 53201

Contact Person: Allen Schuh,
Manager, Safety and Regulatory Engineering
Telephone: 414-647-4385; Fax: 414-647-4090

Date Prepared: December 11, 2001
2. **Device Name:** GE LOGIQ 5 Diagnostic Ultrasound System
Ultrasonic Pulsed Echo Imaging System, 21 CFR 892.1560, 90-IYO
3. **Marketed Device:** GE LOGIQ 7 diagnostic ultrasound system, 510(k) Numbers K010329 currently in commercial distribution.
4. **Device Description:** The GE LOGIQ 5 is a full featured general purpose diagnostic ultrasound system. It consists of a mobile console approximately 52 cm wide, 99 cm deep and 135 cm high that provides digital acquisition, processing and display capability. The user interface includes a computer keyboard, specialized controls, color video CRT display and LCD touch screen. The modification makes a system with numerous high quality features available to diagnostic ultrasound users in a lower market segment.
5. **Indications for Use:** The device is intended for use by a qualified physician for ultrasound evaluation of Fetal; Abdominal; Pediatric; Small Organ (breast, testes, thyroid); Neonatal Cephalic; Adult Cephalic; Cardiac (adult and pediatric); Peripheral Vascular; Musculo-skeletal Conventional and Superficial; Urology (including prostate); Transesophageal; Transrectal; Transvaginal; and Intraoperative (abdominal, thoracic, vascular and neurological).
6. **Comparison with Predicate Device:** The GE LOGIQ 5 is of a comparable type and substantially equivalent to the current GE LOGIQ 7. It has the same technological characteristics, is comparable in key safety and effectiveness features, it utilizes similar design, construction, and materials, and has the same intended uses and basic operating modes as the predicate device.

Section b):

1. **Non-clinical Tests:** The device has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical and mechanical safety, and has been found to conform with applicable medical device safety standards.
2. **Clinical Tests:** None required.
3. **Conclusion:** Intended uses and other key features are consistent with traditional clinical practice, FDA guidelines, and established methods of patient examination. The design and development process of the manufacturer conforms with 21 CFR 820, ISO 9001 and EN 46001 quality systems. The device conforms to applicable medical device safety standards and compliance is verified through independent evaluation with ongoing factory surveillance. Diagnostic ultrasound has accumulated a long history of safe and effective performance. Therefore, it is the opinion of GE Medical Systems that the GE LOGIQ 5 Diagnostic Ultrasound is substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JAN 1 0 2002

Mr. Allen Schuh
Manager, GE Ultrasound Safety
and Regulatory Engineering
GE Medical Systems
P.O. Box 414
MILWAUKEE WI 53201

Re: K014097

Trade Name: GE LOGIQ 5 Diagnostic Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: 90 IYN
Regulation Number: 21 CFR 892.1560
Regulation Name: Ultrasonic pulsed echo imaging system
Regulatory Class: II
Product Code: 90 IYO
Dated: December 11, 2001
Received: December 12, 2001

Dear Mr. Schuh:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the GE LOGIQ 5 Diagnostic System, as described in your premarket notification:

Transducer Model Number

3C
5C
E8C

10L
12L
3S
7S
6T
P6D

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801, please contact the Office of Compliance at (301) 594-4591. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or at (301) 443-6597 or at its Internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

If you have any questions regarding the content of this letter, please contact Rodrigo C. Perez at (301) 594-1212.

Sincerely yours,



Nancy C. Brogdon
Director, Division of Reproductive,
Abdominal and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure(s)

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 Ultrasound System

K014097

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P		
Abdominal ^[1]	P	P	P	P	P	P	P	P	P		
Pediatric	P	P	P	P	P	P	P	P	P		
Small Organ ^[2]	P	P	P		P		P	P	P		
Neonatal Cephalic	P	P	P	P	P	P	P	P	P		
Adult Cephalic	P	P	P	P	P	P	P	P	P		
Cardiac ^[3]	P	P	P	P	P	P	P	P	P		
Peripheral Vascular	P	P	P	P	P		P	P	P		
Musculo-skeletal Conventional	P	P	P		P		P	P	P		
Musculo-skeletal Superficial	P	P	P		P		P	P	P		
Other ^[4]	P	P	P	P	P	P	P	P	P		
Exam Type, Means of Access											
Transesophageal	P	P	P	P	P	P	P	P	P		
Transrectal	P	P	P		P		P	P			
Transvaginal	P	P	P		P		P	P			
Transurethral											
Intraoperative ^[5]	P	P	P		P		P	P			
Intraoperative Neurological	P	P	P		P		P	P			
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes renal, GYN/Pelvic

[2] Small organ includes breast, testes, thyroid.

[3] Cardiac is Adult and Pediatric.

[4] Other use includes Urology/Prostate

[5] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV).

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

Nancy C. Brogdon
(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number K014097

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with 3C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P		P		P	P	P		
Abdominal ^[1]	P	P	P		P		P	P	P		
Pediatric	P	P	P		P		P	P	P		
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P		P	P	P		
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN;

[4] Other use includes Urology;

[*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

Nancy C Brogdon
 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number *K014097*

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with 5C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P		P		P	P			
Abdominal ^[1]	P	P	P		P		P	P			
Pediatric	P	P	P		P		P	P			
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P		P	P			
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN/Pelvic, Renal and Aorta-iliac artery;

[4] Other includes urology

[*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal,
 and Radiological Devices

510(k) Number

Maureen C. Brogdon
K014097

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with E8C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P		P		P	P			
Abdominal ^[1]	P	P	P		P		P	P			
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P		P	P			
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal	P	P	P		P		P	P			
Transvaginal	P	P	P		P		P	P			
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN/Pelvic;

[4] Other use includes Urology/Prostate;

[*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

Nancy C Brogdon
 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K014097

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with 10L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P		P		P	P			
Abdominal											
Pediatric	P	P	P		P		P	P			
Small Organ ^[2]	P	P	P		P		P	P			
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P		P	P			
Musculo-skeletal Conventional	P	P	P		P		P	P			
Musculo-skeletal Superficial	P	P	P		P		P	P			
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative ^[5]	P	P	P		P		P	P			
Intraoperative Neurological	P	P	P		P		P	P			
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E


Notes: [2] Small organ includes breast, testes, thyroid.

[5] Intraoperative includes abdominal, thoracic, and vascular.

[*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number K014097

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with 12L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric	P	P	P		P		P	P	P		
Small Organ ^[2]	P	P	P		P		P	P	P		
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P		P	P	P		
Musculo-skeletal Conventional	P	P	P		P		P	P	P		
Musculo-skeletal Superficial	P	P	P		P		P	P	P		
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative ^[5] (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [2] Small organ includes breast, testes, thyroid.

[*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal,
 and Radiological Devices

510(k) Number

K014097

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with 3S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P		
Abdominal ^[1]	P	P	P	P	P	P	P	P	P		
Pediatric	P	P	P	P	P	P	P	P	P		
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic	P	P	P	P	P	P	P	P	P		
Cardiac ^[3]	P	P	P	P	P	P	P	P	P		
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P	P	P	P	P	P	P		
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN;

[3] Cardiac is Adult and Pediatric;

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

Nancy C Brogdon
(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number *K014097*

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with 7S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P		
Abdominal ^[1]	P	P	P	P	P	P	P	P	P		
Pediatric	P	P	P	P	P	P	P	P	P		
Small Organ (specify)											
Neonatal Cephalic	P	P	P	P	P	P	P	P	P		
Adult Cephalic	P	P	P	P	P	P	P	P	P		
Cardiac ^[3]	P	P	P	P	P	P	P	P	P		
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN;

[3] Cardiac is Adult and Pediatric.

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

Nancy C Brogdon
 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number 6014097

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with 6T Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac	P	P	P	P	P	P	P	P	P		
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal	P	P	P	P	P	P	P	P	P		
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

Nancy Brogdon
 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number *HD14097*

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ 5 with P6D Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac ^[3]				P							
Peripheral Vascular				P							
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [3] Cardiac is Adult and Pediatric.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription User (Per 21 CFR 801.109)

Nancy C. Brody
 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number RD14097